

M A N U A L R E V I S I O N

The instruments listed in Table 1 have been modified to increase their frequency range from 100 kc to 500 kc. Specifications for these instruments are:

Frequency range: DC to 500 kc
Resolution: 2 microseconds

TABLE 1

Model No.	Serial Numbers Affected
7050	322, 367, 422, 457, 459, 473, 478, 480, 482 and above
7150	1955 and above
7151	519 and above
7250	582, 583, 586 through 596, 599 and above
7350	1837 and above
7351	317, 336, 341, 343 and above
8050	108 and above
8150	131 and above
8151	135 and above
8250	104 and above
8350	120 and above
8351	113 and above

Instruments listed in Table 2 have been modified to increase their frequency range from 1 mc to 2 mc. Specifications for these instruments are:

Frequency range: DC to 2 mc
Resolution: 0,5 microseconds

TABLE 2

Model No.	Serial Numbers Affected
7060	752 and above
7160	1359 and above
7260	462 and above
7360	5276 and above
8060	110, 111, 114, 117 and above
8160	140 and above
8260	102 and above
8360	171 and above

MR 51 (4-63)

(Beckman/Berkeley, Manual 7260, 1959)